



PUSHPAGIRI
COLLEGE OF DENTAL SCIENCES
NAAC ACCREDITED



Biomedical Department

Complaint record 2021-22

Dr. Aby Mathew T. MDS
Principal
Pushpagiri College of Dental Sciences



Comp	source	Complaint Type	Name	Phone
168	CONSERVATIVE PG	chair water leak	salooja	
167	CONSERVATIVE PG	micromotor not working	salooja	873
166	CONSERVATIVE PG	Airotor water leaking	Salooja	403
165	PERIO UG	Movement Complaint	JOSNA	268
164	ORTHO PG	chair up and down movement not working	SAVITHA	0276
163	Priority Clinic	3way syringe not working	Suvarna	
162	PEDO PG	xray viewer not working	Sheela	
161	PERIO UG	scalar not working	Josna	
160	PEDO PG	xray viewer not working	sheela	
159	PEDO PG	chair up and down movement not working	sheela	
158	PEDO PG	3way syringe not working	sheela	
157	PROSTHO PG	suction not working	SAVITHA	
156	PERIO UG	water not coming in scalar	josina	0230
155	PERIO UG	water not coming in scalar	josina	0229
154	RADIOLOGY	Error 06	Ajitha	
153	PEDO PG	Not working	Sheela	
152	ORTHO PG	air leak	Rajani	
151	PEDO PG	light on/off switch complaint	sheela	
150	PERIO PG	scalar not working	josina	0265
149	RADIOLOGY	light on/off switch complaint	ajitha	
147	PERIO UG	chair up and down movement not working	josina	0268
146	CONSERVATIVE UG	Not working	salooja	
145	CONSERVATIVE UG	Not working	salooja	
144	ORTHO PG	air leak	rajini	
143	PEDO UG	3way syringe not working	sheela	
142	PEDO UG	water not coming in airotor	sheela	
141	PEDO UG	3way syringe not working	sheela	
140	PEDO PG	chair light not working	sheela	
139	PEDO UG	suction not working	sheela	
138	PEDO UG	Movement Complaint	sheela	
137	PEDO UG	water not coming in airotor	sheela	
136	ORTHO PG		rajini	
135	ORTHO UG		rajini	
134	ORTHO PG		rajini	
133	PROSTHO PG	lab micromotor unit complaint	SAVITHA	0
132	PROSTHO CERAMIC	Not working	SAVITHA	
131	PROSTHO CERAMIC	Not working	SAVITHA	-1
130	PROSTHO CERAMIC	air leak	SAVITHA	711
129	CONSERVATIVE UG	suction not working	salooja	366
128	CONSERVATIVE UG	water not coming in airotor	salooja	373
127	CONSERVATIVE UG	suction low pressure	salooja	374
126	CONSERVATIVE UG	airotor low power	salooja	362
125	CONSERVATIVE UG	water not coming in airotor	salooja	368
124	CONSERVATIVE UG	suction not working	salooja	364
123	CONSERVATIVE UG	water not coming in airotor	salooja	364

122	ORTHO PG	scalar not working	rajini	
121	OMFS PG	A03 ERROR	MARY JIJI	
120	OMR	Not working	LEELAMMA	
119	OMFS PG	micromotor speed controler complaint	maryjiji	00078
118	OMFS PG	foot control not working	maryjiji	00077
117	PROSTHO PG	micromotor not working	savitha	
116	PROSTHO PG	water not coming in airtor	savitha	
115	PROSTHO UG	micromotor not working	savitha	
114	CONSERVATIVE UG	3way syringe not working	Salooja	946
113	CONSERVATIVE UG	chair light not working	Salooja	368
112	CONSERVATIVE UG	3way syringe not working	Salooja	376
111	CONSERVATIVE UG	3way syringe not working	Salooja	374
110	CONSERVATIVE UG	chair up and down movement not working	Salooja	365
109	ORALPATHOLOGY	Not working	Susan	600
108	CONSERVATIVE PG	light on/off switch complaint	salooja	
107	PROSTHO PG	3way syringe not working	savitha	0643
106	PROSTHO PG	micromotor not working	savitha	640
105	PROSTHO UG	model trimmer water leaking	savitha	
104	ORTHO PG	Not working	sabitha	
103	CONSERVATIVE UG	water not coming in airtor	salooja	
102	PERIO PG	suction not working	josina	0262
101	PERIO UG	suction not working	josina	0230
100	PERIO UG	chair up and down movement not working	josina	0225
99	PERIO UG	chair up and down movement not working	josina	0228
98	PROSTHO UG	chair up and down movement not working	savitha	0613
97	PROSTHO UG	chair up and down movement not working	savitha	622
96	PROSTHO UG	backrest up and down movement not working	savitha	0621
95	PROSTHO UG	chair up and down movement not working	savitha	0617
94	PROSTHO UG	chair up and down movement not working	savitha	0616
93	OMFS PG	chair light not working	maryjiji	0074
92	OMFS PG	foot control not working	maryjiji	0071
91	CONSERVATIVE UG	Not working	Salooja	
90	PERIO PG	Not working	Josna	
89	PERIO UG	Not working	Josna	
88	ORTHO PG	chair up and down movement not working	sabitha	
87	ORTHO PG	backrest up and down movement not working	sabitha	
86	ORTHO UG	pressure port complaint	sabitha	
85	ORTHO UG		sabitha	
84	OMFS PG	chair up and down movement not working	maryjiji	0074
83	ORTHO PG	Not working	sabitha	
82	ORTHO UG	micromotor not working	sabitha	
81	CONSERVATIVE PG		salooja	
80	OMFS UG	backrest up and down movement not working	maryjiji	000
79	CONSERVATIVE UG	booster on/off switch complaint	salooja	
78	CONSERVATIVE PG	micromotor not working	salooja	
77	ORTHO PG	scalar not working	rajini	
76	ORTHO UG	Not working	rajini	

75	ORTHO PG	airotor switch complaint	sabitha	
74	CONSERVATIVE UG	machine not getting ON	salooja nadarajan	
73	PERIO UG	A03 ERROR	Josina	
72	PROSTHO PG	3way syringe not working	savitha	
71	PROSTHO PG	3way syringe not working	savitha	
70	PEDO PG	air leak	sheela	
69	PERIO PG	suction not working	josina	
68	PERIO UG	suction not working	josina	
67	PERIO UG	suction not working	josina	
66	PERIO PG	suction not working	josina	
65	PERIO PG	suction not working	josina	
63	OMFS PG	micromotor not working	mary jiji	0077
62	OMFS UG	Not working	mary jiji	
61	ORTHO PG	suction low pressure	Rajani	356
60	OMFS UG	Spittoon block	mary jiji	
59	PERIO UG	chair up and down movement not working	josina	
58	PERIO UG	chair up and down movement not working	josina	
57	PEDO PG	scalar not working	sheela	
56	OMFS UG	chair up and down movement not working	mary jiji	40
55	PEDO PG	water not coming in airotor	sheela	-1
54	PERIO PG	scalar not working	josina	
53	OMFS PG	chair up and down movement not working	mary jiji	0072
52	ORTHO PG	chair up and down movement not working	sabitha	
51	PROSTHO PG	Not working	savitha	
50	PROSTHO PG	micromotor not working	savitha	
49	PERIO PG	suction not working	josina	
48	OMFS PG	backrest up and down movement not working	mary jiji	-1
47	CONSERVATIVE PG	machine not getting ON	salooj	
46	ORTHO UG	chair up and down movement not working	sabitha	
45	ORTHO PG	suction not working	sabitha	
44	ORTHO PG	suction not working	sabitha	
43	ORTHO UG	chair up and down movement not working	sabitha	
42	ORTHO UG	water not coming in spittoon	sabitha	
41	CONSERVATIVE PG	chair water leak	salooja	
40	PROSTHO UG	backrest up and down movement not working	savitha	
39	OMFS UG	chair up and down movement not working	mary jiji	
38	PROSTHO PG	micromotor connector complaint	savitha	
37	CONSERVATIVE PG	machine not getting ON	salooja nadarajan	
36	ORTHO PG	suction not working	sabitha	
35	ORTHO UG	chair up and down movement not working	sabitha	
34	PERIO UG	scalar not working	josina	
33	PERIO PG	chair light not working	josina	
32	PEDO PG	scalar not working	sheela	
31	PEDO PG	air leak	sheela	
30	PROSTHO UG	chair side swich not working	savitha	
29	PROSTHO PG	machine not getting ON	savitha	
28	OMR	Movement Complaint	ajitha	

27	ORTHO PG	Not working	sabitha	
26	OMFS PG	Not working	mary jiji	
25	PERIO UG	Not working	josina	
24	PROSTHO PG	backrest up and down movement not working	savitha	
23	CONSERVATIVE PG	spitton leakage	salooja	
22	CONSERVATIVE PG	micromotor not working	salooj	
21	PROSTHO PG	spitton leakage	savitha	
20	PROSTHO PG	lathe machine not working	savitha	
19	CONSERVATIVE PG	light on/off switch complaint	salooja	
18	PROSTHO UG	micromotor connector complaint	savitha	
17	CONSERVATIVE PG	water not coming in spittoon	salooj	
16	PROSTHO UG	chair up and down movement not working	savitha	
15	PROSTHO UG	chair light not working	savitha	
14	PROSTHO UG	micromotor not working	savitha	
13	PROSTHO UG	chair up and down movement not working	savitha	
12	PERIO PG	suction not working	josina	
11	ORTHO UG	micromotor not working	rajini	
10	ORTHO PG	light on/off switch complaint	sabitha	
8	PEDO PG		sheela	
7	PEDO PG	suction not working	sheela	
4	IMPLANT CLINIC	suction not working	vinitha	

Pushpagiri College of Dental Sciences

Complained D	Description	Action Taken	Assigned
31-08-2022		Retraction valve c	Beegam Fathima
31-08-2022	Need to change the micromotor		
31-08-2022		REPAIRED	Beegam Fathima
31-08-2022	NEED TO CHANGE THE SIDE S	REPAIRED	Manu Easo George
30-08-2022		REPAIRED	Manu Easo George
29-08-2022		3 way syringe and	Biju George
26-08-2022		Xray viewer repla	Biju George
26-05-2022	Micromotor- Scaler pcb com	New pcb fixed	Manu Easo George
23-08-2022		REPAIRED	Manu Easo George
23-08-2022		REPAIRED	Biju George
23-08-2022		REPAIRED	Biju George
20-08-2022		Transformer chan	Biju George
23-08-2022		Booter bottle tub	Manu Easo george
23-08-2022		Booster bottle tut	Manu Easo george
23-08-2022	X ray tube complaint. Need to be changed Und		Beegam Fathima
23-08-2022	One repaired		
23-08-2022		Tube changed	Manu Easo george
23-08-2022		Switch changed	Beegam Fathima
22-08-2022		REPAIRED	Manu Easo George
20-08-2022		REPAIRED	Manu Easo George
20-08-2022		REPAIRED	Manu Easo George
19-08-2022	Being Tulip B-100	New adapter give	Beegam Fathima
19-08-2022	Beecool	New adapter give	Beegam Fathima
18-08-2022		Tube changed	Beegam Fathima
18-08-2022		3 way syringe cha	Beegam Fathima
18-08-2022		Tube changed	Biju George
18-08-2022		3 way syringe cha	Biju George
18-08-2022	switch complaint		
18-08-2022		REPAIRED	Biju George
18-08-2022		REPAIRED	Beegam Fathima
18-08-2022		REPAIRED	Biju George
18-08-2022		REPAIRED	Biju George
18-08-2022		REPAIRED	Biju George
18-08-2022	bur broken. pls taken to store	REPAIRED	Beegam Fathima
18-08-2022	handpiece fell down and bur	REPAIRED	Beegam Fathima
17-08-2022		REPAIRED	Biju George
17-08-2022		REPAIRED	Biju George
17-08-2022	FOOT SWITCH NOT WORKING	Foot control switc	Biju George
17-08-2022		Suction pcb and p	Biju George
17-08-2022		REPAIRED	Beegam Fathima
17-08-2022		REPAIRED	Beegam Fathima
17-08-2022		REPAIRED	Beegam Fathima
16-08-2022		REPAIRED	Beegam Fathima
16-08-2022		suction hose chan	Biju George
16-08-2022		REPAIRED	Beegam Fathima

17-08-2022		Tube changed	Biju George
16-08-2022		REPAIRED	Manu Easo George
16-08-2022		REPAIRED	Manu Easo george
13-08-2022			
13-08-2022			Biju George
13-08-2022		REPAIRED	Manu Easo george
13-08-2022		REPAIRED	Manu Easo george
13-08-2022		REPAIRED	Manu Easo george
11-08-2022		Repaired	Fathima
11-08-2022	3way not working	Repaired 3 way, b	Biju George
11-08-2022	Airotor not working	3way syringe and	Bijuv, Fathima
11-08-2022		3way syringe char	Beegam Fathima
11-08-2022		Side switch and 4	Biju George
12-08-2022		Bulb changed	Beegam Fathima
12-08-2022		REPAIRED	Biju George
11-08-2022		REPAIRED	Biju George
11-08-2022		REPAIRED	Biju George
11-08-2022		Tube changed	Manu Easo george
11-08-2022		1	
11-08-2022	chair no 2-speed control not working,chair no 4-hand piece not working,cha		
10-08-2022			
10-08-2022		REPAIRED	Manu Easo George
10-08-2022			
10-08-2022		REPAIRED	Beegam Fathima
10-08-2022		Repaired	Biju george
10-08-2022		Repaired, side swi	Biju george, Beegam fathim
10-08-2022		Side switch changed	
10-08-2022		Repaired	Biju george
10-08-2022		Side switch chang	Beegam fathima
09-08-2022		Repaired	Biju George
09-08-2022		Repaired	Biju George
06-08-2022		Tube changed	Beegam Fathima
06-08-2022		Tube changed	Beegam fathima
05-08-2022	3 uv cabinets are complaint	Repaired 2 numb	Beegam Fathima
04-08-2022		1 Repaired	Biju George
03-08-2022		Repaired	Biju George
03-08-2022			
02-08-2022	REPLACE THE WATER TUBING	Repaired	Beegam Fathima
02-08-2022			
30-07-2022	CONNECTION PROBLEM	Repaired	Biju George
30-07-2022	FIXED THE ISSUE	Repaired	Biju George
30-07-2022	x smart ENDOMOTOR. NEED	PENDING	
29-07-2022	UP/DOWN MOTOR REPLACED	Repaired	Biju George
29-07-2022			
29-07-2022	xray char up down not worki	Repaired	Biju George
26-07-2022	RE FIX THE ISSUE	Repaired	
26-07-2022		Repaired	

26-07-2022	2 chair	Repaired	
23-07-2022	Plug top burned	PENDING	Beegam Fathima
23-07-2022	Drainage water tank was blo	Repaired	Beegam Fathima
23-07-2022	CHAIR NO 6	Repaired	
23-07-2022		Repaired	
23-07-2022	Pneumatic suction was on	Repaired	Beegam Fathima
22-07-2022		Repaired	Beegam Fathima
21-07-2022	Capacitor complaint	capacitopr change	Beegam Fathima
21-07-2022	Suction tube block	Repaired	Beegam Fathima
20-07-2022		REPAIRED	Manu Easo George
20-07-2022		REPAIRED	Biju George Abraham
20-07-2022		New micromotor	Beegam Fathima
19-07-2022	Can we use for inspection pu	Beyond repair, Sp	Biju George
19-07-2022	Spittoon block	Repaired	Beegam Fathima
19-07-2022	chair no 39	Repaired	Beegam Fathima
18-07-2022		Repaired	Biju George
18-07-2022		Repaired	Biju George
18-07-2022		Repaired	Biju George
18-07-2022		Side switch cable	Biju George
16-07-2022		Repaired	Biju George
13-07-2022	RE-FIX THE PLUG PROPERLY &	Repaired	
12-07-2022		Repaired	Biju George
08-07-2022	1 dental chair	UP/Down Motor	Biju George
08-07-2022			
07-07-2022	DENTAL CHAIR NOT WORKING		
06-07-2022			
06-07-2022			
05-07-2022	X SMART NOT WORKING	Timer new fixed	Biju George
05-07-2022	9 dental chair		
04-07-2022		Repaired	Biju George
01-07-2022		Repaired	Biju George
01-07-2022			
01-07-2022			
01-07-2022		3 Way syringe cha	Beegam Fathima
01-07-2022			
01-07-2022		Repaired	Biju George
30-06-2022		new connector giv	Biju George
29-06-2022		Coil complaint to	Biju George
28-06-2022			
28-06-2022			
28-06-2022		REPAIRED	Biju George Abraham
28-06-2022		REPAIRED	Biju George Abraham
28-06-2022	130- 131-	wood pecker scale	Beegam Fathima
28-06-2022		Repaired	Biju George
28-06-2022			
25-06-2022			
25-06-2022		Repaired	Biju George

25-06-2022		PCB REPAIRED	Biju George
24-06-2022	OMFS PG 78	Bulb changed	Beegam Fathima
24-06-2022		REPAIRED	Beegam Fathima
23-06-2022	prostho pg- 647 Main board	Main Board change	Biju George
23-06-2022			
23-06-2022			
11-06-2022			
08-06-2022	MODEL TRIMMER NOT WORKING	Model trimmer arranged	St;Geo
04-06-2022		LIGHT SWITCH CHANGED	Biju
04-06-2022	Chair no 1 AIR LEAKING .CHAIR NO 6.HANDPIECE LEAKING.		
02-06-2022			
02-06-2022	light complaint	Repaired	Biju
02-06-2022		Bulb changed	Beegam Fathima
02-06-2022		on Action Taken	
02-06-2022		Repaired	Biju
28-05-2022		Repaired	Biju
09-05-2022		Repaired	Biju
09-05-2022			
07-05-2022		Repaired	Beegam Fathima
07-05-2022	Model trimmer was working	Instructed to connect	Beegam Fathima
30-04-2022	Repaired	capacitor change	Biju

Device Type	AssetID	Part Change	To
DENTAL CHAIR		NRV valve	biomedical
DENTAL CHAIR	873		BIOMEDICAL
DENTAL CHAIR	403		BIOMEDICAL
DENTAL CHAIR	268		BIOMEDICAL
DENTAL CHAIR	0276		BIOMEDICAL
DENTAL CHAIR		3 way syringe and tubing's	BIOMEDICAL
DENTAL CHAIR		Xray viewer	BIOMEDICAL
DENTAL CHAIR		Micromotor - Scaler PCB	BIOMEDICAL
		ON/OFF SWITCH REPLACE	biomedical
DENTAL CHAIR		SIDE SWITCH REPLACED	biomedical
DENTAL CHAIR		3 WAY SYRINGE REPLACED	biomedical
DENTAL CHAIR		Transformer	BIOMEDICAL
SCALER UNIT	0230	4 mm tube	biomedical
SCALER UNIT	0229	4 mm tube	biomedical
CBCT			BIOMEDICAL
Welder			BIOMEDICAL
DENTAL CHAIR		6 mm Tube	BIOMEDICAL
DENTAL CHAIR		On/ Off switch	biomedical
SCALER UNIT	0265		biomedical
X RAY MACHINE		ON/OFF SWITCH REPLACE	biomedical
	0268		biomedical
LIGHT CURE UNIT		5V adapter	BIOMEDICAL
LIGHT CURE UNIT		9 V adapter	BIOMEDICAL
		6 mm tube changed	biomedical
DENTAL CHAIR		3 way syringe	biomedical
DENTAL CHAIR		4mm tube	biomedical
DENTAL CHAIR		3 way syringe	biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
SCALER UNIT			biomedical
LAB MICROMOTOR			biomedical
LAB MICROMOTOR			biomedical
LAB MICROMOTOR	0		BIOMEDICAL
circular saw			BIOMEDICAL
Vibrator	-1		BIOMEDICAL
SAND BLASTER	711	Foot control switch, 6 mm	BIOMEDICAL
DENTAL CHAIR	366	suction pcb, Plug top	BIOMEDICAL
DENTAL CHAIR	373		BIOMEDICAL
DENTAL CHAIR	374		BIOMEDICAL
DENTAL CHAIR	362		BIOMEDICAL
DENTAL CHAIR	368		BIOMEDICAL
DENTAL CHAIR	364	1" hose tube	BIOMEDICAL
DENTAL CHAIR	364		BIOMEDICAL

DENTAL CHAIR		6mm Tube changed	biomedical
AUTOCLAVE - B CLASS			BIOMEDICAL
DENTAL CHAIR			BIOMEDICAL
DENTAL CHAIR	00078		biomedical dept
DENTAL CHAIR	00077	Foot control switch	biomedical dept
LAB MICROMOTOR			biomedical
AIROTOR HANDPIECE			biomedical
LAB MICROMOTOR			biomedical
DENTAL CHAIR	946	3way syringe	Biomedical
DENTAL CHAIR	368	Bulb 12v 55	Biomedical
DENTAL CHAIR	376	3 way syringe, 4 mm conn	Biomedical
DENTAL CHAIR	374	3 way syringe	Biomedical
DENTAL CHAIR	365	Side switch, 4 mm tube	Biomedical
MICROSCOPE	600	Bulb 6v 20 w	Biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR	0643		biomedical
DENTAL CHAIR	640		biomedical
MODEL TRIMMER- SINGLE DISC		8 MM & 6 MM tube chang	biomedical
AIROTOR HANDPIECE			biomedical
DENTAL CHAIR			biomedical
	0262		biomedical
	0230		biomedical
	0225		biomedical
	0228		biomedical
DENTAL CHAIR	0613		biomedical
DENTAL CHAIR	622	Side switch	biomedical
DENTAL CHAIR	0621	Side switch	biomedical
DENTAL CHAIR	0617		biomedical
DENTAL CHAIR	0616	Side switch	biomedical
DENTAL CHAIR	0074		
DENTAL CHAIR	0071	foot control unit	
UV cabinet		UV tube	Biomedical
UV cabinet		Uv tube	Biomedical
UV cabinet		3 uv tube	Biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
MODEL TRIMMER- SINGLE DISC		TUBE - CHESA SUCTION TU	biomedical
DENTAL CHAIR	0074		
AIROTOR HANDPIECE			biomedical
LAB MICROMOTOR			biomedical
Electro cautery unit			biomedical
DENTAL CHAIR	000	UP/DOWN Motor CHANGED	
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
LAB MICROMOTOR		SWITCH REPLACED	biomedical

DENTAL CHAIR			biomedical
DENTAL CHAIR			BIOMEDICAL
AUTOCLAVE - B CLASS			BIOMEDICAL
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR		8 mF Capacitor	biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
			biomedical
DENTAL CHAIR	0077	micromotor connector	biomedical
ULTRASONIC CLEANER			biomedical
DENTAL CHAIR	356		BIOMEDICAL
DENTAL CHAIR			biomedical
DENTAL CHAIR		Light switch cover change	biomedical
DENTAL CHAIR		Suction cavity holder change	biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR	40		biomedical
AIRATOR HANDPIECE	-1		biomedical
SCALER UNIT			biomedical
DENTAL CHAIR	0072		biomedical
DENTAL CHAIR		UP/DOWN Motor	biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
SUCTION UNIT			biomedical
DENTAL CHAIR	-1		biomedical
X RAY MACHINE		X ray timer	BIOMEDICAL
DENTAL CHAIR			biomedical
SUCTION UNIT			biomedical
SUCTION UNIT			biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR		3 way syringe	
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
MICROMOTOR HAND PIECE		micromotor connector	biomedical
AUTOCLAVE - B CLASS			
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
SCALER UNIT			biomedical
		Bulb	biomedical
DENTAL CHAIR		S19A0032P(131), S19A011	biomedical
DENTAL CHAIR		6mm tube 2 m	biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical

DENTAL CHAIR			biomedical
DENTAL CHAIR		Bulb 12v 50w	biomedical
SUCTION UNIT			biomedical
DENTAL CHAIR		Main PCB -Onyx	biomedical
DENTAL CHAIR			
DENTAL CHAIR			biomedical
LATHE MACHINE			biomedical
DENTAL CHAIR		Light Switch	biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR			biomedical
DENTAL CHAIR		Bulb	biomedical
LAB MICROMOTOR			biomedical
DENTAL CHAIR			biomedical
SUCTION UNIT			BIOMEDICAL
LAB MICROMOTOR			bio medical
DENTAL CHAIR			bio medical
DENTAL CHAIR			biomedical
MODEL TRIMMER- SINGLE DISC			biomedical
SUCTION UNIT			biomedical

13/10/2021	PG	C7800121	Handpiece ^(Right side down) Not working		
15/10/21	PG	PU	Light case and caping ^(Right side down)	Repair	Biji
22/11/21	PG	C353	Chair Movement Not working		
25/11/21	PG	Chair no: 355	Light case not working	Board Complaint	Replaced
29/11/21	UG	Chair no:	Chair Movement Not working	repaired.	Miya Biji
29/11/21	PG	PU	Model handles one side get heated On training		
7/12/21	PG		Light chair light not working	Switch need to be replaced.	
8/12/21	UG	332 micro motor	Not working	Early send it to some department	Uli
13/12/21	PG	Acylizer	Acylizer Not working		
31/12/21	PG	358 Section	Section Complaint	Repaired	Jithin 31/12/2021
31/12/21	PG	353	Scaler Complaint		
6/1/2022	PG	OBE 9170 C8100039 C8890086	Handpiece ③ Not Working	Repaired	Jithin
25/1/2022	UG	Dental Lath	Lath Dental Lath Not Working		
15/2/2022	PG	Autoclave	Drain water connector tighted.	Repaired	Jithin 15/2/2022
25/2/22	UG	Micro motor	Note working (1945, 196)	Repaired	Jithin
"	UG	Lathe	Buff Not working		
11/3/22	UG	Micro-motor for Handpiece	Note working (Hand over to Biji)		

18/3/22	Py 2nd Chair	light not working.	
18/3/22	Autoclave	Wabe complaint	
21/3/22	Ortho Pn	Chair no: 355	Light Not Working
8/4/22	Pu	Sand Blaster	not working
"	u6	Chair	not working
18/4/22	u6	micromotor (0347) HandPic	not working (1 and over Fatma)
18/4/22	Pu	Chair No: 108	Light complaint
30/4/22	Pu7	Chair no: 353	Light complaint Chair not working
13/5/22	P9	holder (2)	Not working
19/5/22	P6 - 858	Denial chair	light arm broken
20/5/22	u6	Milux motor Hand	not working.
24/5/22	361	Complains Scales Dental Chair Complains	Scales complaint
31/5/22	346	Chair no: 346	broken light
7/6/22	Pu	Chair no: 345	Scales not working
21/6/2022	u6	Micromotor	Not working (0347)

	Board complaint		
	Repaired		F. Thami
	PCB complaint		
	Repaired		F. Thami 11/4/22
	in case		Returned F. Thami
	Repaired		F. Thami 20/4/22
	Repaired board changed (min)		F. Thami 20/4/22
	tip changed working		F. Thami 15/5/22
	Rotating arm repaired		F. Thami 14/5/22
	changed (re)		F. Thami 18/5/22
	Repaired		F. Thami 22/4/22
			Bida
			Bida
			Bida

2/16/22		Dental chair	
2/10/22	Ortho PG	Dental chair 2/19/22	Three ways not working
4/10/22	Ortho PG	Dental chair 4/19/22	Suction not working
9/2/22	Ortho PG	Dental chair	chair no. 355 up/down movement not working
9/27/22	Ortho PG	Dental chair	have dental chair
25/3/22	Ortho PG	Micromotor with Handpiece	not working
20/7/22	Ortho PG	Lab micromotor complaint	not working
°	Ortho PG	Amblets not working	not working
2/8/22	Ortho PG	Model former	water leaking
3/8/22	Ortho PG	WELDER MISTON	Footstep broken
	Ortho PG	DIAMOND GRIND	not working
	Ortho PG	"	not working
4/8/22	Ortho PG	Dental chair	light of movement not working
6/8/22	PG	Autoclave Penetrive	complaint (not working)
17/8/22	PG	Chair	Scalce not working
18/8/22	PG	Chair	Air leaking
18/8/22	PG	Micromotor	Micromotor complaint
18/8/22	UC	Micromotor	Not working

Repaired to 2/17/22
Up/down movement replaced installed and working
Repaired Switch changed
Tube changed
Beyond repair, Need to replace it spare parts ordered
new Penetrive given
Repaired 16/8/22

John
2/17/22

John
5/10/22

John
5/13/22

John
5/13/22

John
5/13/22

John
5/13/22

226255

28/4/21	PG	101	Air leaking Chair not working.		Bye 29/4/21
28/4/21	PG	Model hammer Double disc	Motor Complaint	Sent for repair to Mr. Ajib (Thursday)	at home 29/4/21
22/6/21	PG	Chair-103.	low Suction cap mixing	used Suction cap.	Bye 22/6/21
	PG	Chair-101	air leak Not working	Repaired.	Bye 22/6/21
	PG	Chair 106	Air not getting	repaired	Bye 22/6/21
11/8/21	PG		light cure Not working (2nos) Cone taken for repair. <u>broken</u> , <u>broken</u> <u>broken</u> both found not adequate. <u>broken</u> .	Both light cure repaired and cones to dispatch	Bye
2/9/21	PG	Chair-10102	Water hose for Air for Hand Piece Not working	Sealed	Bye

13/3/21	pu 5-16	rental chair	suction not working
8/3/2021	PG-	Dented chi	Air leaking
10/3/2021	P-G	Model Trime	Not working (Buen)
"	P-G	Chair No: 106	Hand Piece Not working
"	P-G	Chair no: 102	Scale Not Working
17/4/21	P-G	Chair no: 103	Air Leaking
"	"	" 104	
"	"	" 105	
"	"	102	suction Not working
17/4/21	PG	(base)	Motor complaint 9226255
28/4/21	PG	101	Air leaking
"	"	"	chair not working.
23/4/21	PG	Model tanner Double disc	Motor Complaint
22/6/21	PG	chair-103	low suction cap mixing
	PG	chair-101	airlet Not working
	PG	chair 106	Air not getting
17/8/21	PG		light cure Not working (2nos) Send taken for repair <u>khilla</u> , <u>19/8/21</u> <u>khilla</u> both found not adequate <u>Brickend</u> .
2/9/21	PG	chair no: 102	Water hose Air for Hand Piece Not working

Repaired

13/3/21
St. James
 8/3/21
Bun
Bun
Bun
Bun
Bun
 17/4/21

Bun
 28/4/21

Sent for repair
 to Mr. Ajith (Thursday)

St. James
 20/4/21
B. L. K.
 22/6/21

used suction cap.

Repaired.
 repaired

khilla
 22/8/21
khilla
 22/8/21

Both light cure repaired
 and crown to department

khilla

Sent

Bun

11/5/2021 Perio UG wash basin leaking

28/6/2021 Perio UG
[Chair no: 0223] Light not working

16/6/2021 Perio PG.
[Chair no: 0268] Suction not working

23/7/2021 Perio PG.
[Chair no: 2012] Suction not working

30/2/2021 Perio PG.
[Chair no: 0211] water leaking

27/8/2021 Perio PG.
[Chair no: 0266] Scaler unit cable not working

15/9/2021 Perio UG.
[Chair no: 0855] Scaler handpiece complaint.
[Decreased water flow]

29/9/2021 Perio PG New chair movement
complaint.

29/11/2021 Perio UG
[Chair no: 0281] Air leaking.

27/12/2021 Perio UG auto clare machine complaint (Runs)

Repaired Jathini 11/5/21

Bulb changed working Jathini 28/6/21

Motor complaint Jathini 16/6/21

Repaired Biji 23/7/21

Repaired Jathini 30/2/21

Repaired Jathini 27/8/21

Repaired Biji 15/9/21

Side switch changed Re:paired Biji 29/9/21

Repaired Biji

water jella cleaned Re:paired Jathini 27/12/21

13/7/22	Peno PG		Scissors not working	Repaired	F. H. M.
13/7/22	Peno UG	Chair no: 0230	Dental chair not working	Capacitor changed	F. H. M. 23/7/22
	"	Chair no: 0230	Dental chair not working	Working	
	"	Chair no: 0230	Dental chair not working		
23/7/22	Peno pg: 261	Dental chair	Suction not working	Repaired	} F. H. M. 23/7/22
23/7/22	UG	Autoclave	NO ₂ leak error	Repaired	
23/7/22	UG - 231	Dental chair	Scale Booster Bottle leaking	Repaired	
23/7/22	UG - 229	Dental chair	Suction pressure low	Repaired	} F. H. M. 23/7/22
23/7/22	PG - new chair	Dental chair	Side switch complaint	Repaired	
23/7/22	PG - 265	Dental chair	Suction Complaint	Repaired	} F. H. M. 23/7/22
4/8/22	PU - 1	UV cabin	Not working		
4/8/22	UC1 - 3	UV cabin	Not working	UG repaired	
10/8/22	Peno UC1	chair no: 0228	Chair not working		
"	Peno UC1	chair no: 0228	chair not working		
"	Peno UC1	chair no: 0230	Scissors not working	Repaired	Man
"	Peno PU	chair no: 0262	Scissors not working		
12/8/22	PU	269	Water leaking	Repaired	Man
22/8/22	Peno PG	0268	Chair switch complaint	1	
22/8/22	Peno PU	chair no: 0265	Scissors not working	Repaired	F. H. M.

Da Alcha
Da Alcha
Da Alcha
Da Alcha
Da Alcha
Da Alcha
Da Alcha



3/1/2022	Perio. UG	Chair no. (0848)	Scaler board complaint
6/1/2022	Perio PU	Chair no 0263	Suction not working
12/1/2022	Perio PU	Chair no 0269	Suction stand broken
12/1/22	Perio PU	Chair no 0268	Suction not working Scaler water leaking
22/1/22	Perio PU	chair no. 0261	Scaler has no pressure
22/2/22	Perio OG d41	Picazo laser	foot control cable complaint
22/3/22	Perio PU	chair no: 0262	water leaking
		262	Dental chair
		263	Dental chair
		266	Dental chair
		265	Dental chair
		268	Dental chair
23/4/22	UG - 2nd	Dental chair	light complaint
8/6/22	Perio PU	chair no: 0214	Light not working
9/6/22	Perio PU	chair no: 0212	Suction not working
11/6/22	Perio PU	chair no: 0262	Suction not working
14/6/22	Perio UG	suction not working	see
28/6/22	Perio UG	Scaler not working	
	Perio PU	light not working	
2/7/22	Perio PU	chair no:	Suction not working

Repaired

[Signature]
22/1/22

Taken to store for repair

[Signature]
22/6/22

repaired
motor to be changed
Suction bottle to be changed
repaired
repaired

[Signature]
23/4/22

Repaired

Repaired
Switch changed

[Signature]
F. Thami

Repaired
Repaired

[Signature]

DATE	UG/PG - CHAIR NO	ITEM	COMPLAINT	REMARK OF BIODENTAL STAFF	SIGNATURE	REMARK OF DEPT. STAFF
6/08/19	check no 47	Scales	No Water, Air leaking		<i>Fahim</i> 6/8/19	
23/08/19	chair no 46	Suction	not working	Repaired	<i>Fahim</i> 23/8/19	
16/09/19	chair no 46	Leaking	Air Leaking	Repaired	<i>Burn</i> 16/9/19	
22/09/19	chair no 46	Suction	Not working & Damaged	Stand backer		
2/11/19	47	Movement	Movement not working	Switch complaint	<i>Fahim</i> 2/11/19	
4/12/19	49	Dental chair	light complaint	light switch & light cover changed	<i>Fahim</i> 4/12/19	
6/12/19		Autoclave	Manual autoclave washer new	given	<i>Fahim</i>	
23/12/19	47	Analyzer	Analyzer not working	Repaired	<i>Fahim</i> 23/12/19	
25/01/2020	asset id 0191	Dental chair	Movement Complaint	Up/down Motor changed	<i>Fahim</i> 25/1/2020	
1/02/2020	0192	Dental chair	Sample painting done by	IKON	<i>Fathime</i>	
20/2/2020	0192	Dental chair	seat & back cushions fixed		<i>Fathime</i> 20/2/20	
16/1/20	New Dental chair		Light not working	Sensor complaint informed to company (under warranty)	<i>Fathime</i> 16/1/20	
20/1/20	"	Dental chair	-> PCB changed (over clock)			
22/1/20	0193	Dental chair	Not working	Low suction tube and holder changed and repaired	<i>Fathime</i> 22/1/20	
17/3/21	956	Dental chair	Air leaking	Repaired		
23/03/21	337	Dental chair	movement not working	Repaired		
23/03	957	Dental chair	Air leaking	Repaired		
2/10/21	841	"	Light not working	Repaired	<i>Fahim</i> 10/4/21	
9/8/21	084, 189	Dental chair	3Way Not Working	Repaired	<i>Fahim</i> 9/8/21	
20/5/22	Chair 841	Dental chair	Analyzer not working	NAY changed	<i>Fahim</i> 20/5/22	
20/1						

3/6/22	pu	chair	Three way not working
25/6/22	pu	Chair	Foot control not working
"	"	"	"
25/6/22	u4	wood pecker	Scales unit not working
28/6/22	pu	Chair no: 120	Scales not working
"	pu	Chair no: 131	Air leaking Scales not working
6/7/22	pu	puls tester - 2 No's	not working
12/7/22	pu	Disorder	Handpiece no: C5 Z5 6287 not working
18/7/22	pu	scales	not working
18/7/22	pu	chair no: 0112	Three way not working
23/7/22	pu	chair	Suction holder Broken
"	pu	chair	Air leaking
10/8/22	pu	Chair	Suction not working
11/08/22	pu	rotator	Not working

Repaired	
Repaired	
Value to be cleaned, can use water for scaling	
New scales fixed S19A0032P	
Air tube replaced (S New scales fixed (S19A0110P)	
Belt complaint - Taken to store	
not working	
Repair	
Repair	
Need to replace	
Repaired	

1/6/24

J. K. 25/6/22

J. K. 25/6/22

J. K. 28/6/22

J. K. 28/6/22

J. K. 29/7/22

Repair 18/7/22

Repair 18/7/22

J. K. 23/7/22

13/8/22	U-G	Chair 0100	Foot control & side switch not working	Repaired
	U-G	Chair 0101	Hand Pic-coater not Movement Complaint	Repaired
	U-G	Chair 102	Three way & simulator & Section not working	Repaired
	U-G	Chair 099	Side switch not working	Repaired
	U-G	Chair 107	Three way & simulator, foot control not working Spoon - water not working	Repaired A.M.S tube changed
	U-G	Chair 108	Director & foot control not working	
7/8/2022	UG -	830	3way syringe not working	3way syringe repaired
	UG -	831	3way not working	3way syringe repaired
	PGT. 126	Chair	light Complaint	Repaired
	PGT. 129	Chair	light Complaint	switch complaint, to be changed
19/8/22	PGT. 129	Chair	light Complaint	Switch Complaint not working
20/8/22	PGT. 129	Chair	light Complaint	Switch Complaint not working
22/8/22	PGT. 129	Chair	light Complaint	
23/8/22	PGT. 129	"	" not working	
23/8/22	PGT.		welder not working (given to biomedical stores)	

Phil 18/8/2022

John 15/8/2022

John 18/8/22

5/1/2022	Hydroxide		Not working
5/2/2022	PG	chair	side ^{switch} switch Not working
14/2/2022	PU	Light cure	3m GSPE Eclipus Deep cure Not working
15/2/2022	PG - 128	Dental chair	Scaler complaint
15/2/22	PG - 830	Dental chair	3 way syringe all leaking
15/2/22	PG - 130	Dental chair	Scaler Complaint
13/3/22	UG - 11	Dental chair	Diagnos. complaint (3 chairs)
	PU	Chair no 2014	Diagnos not working
18/3/22	U-01	Dental chair	Scalors not working
		Dental chair	Diagnos not working
18/3/23	P-G1	Dental chair 0632	Three way not working
31/3/22	pu	Chair no	Suction Not working

Machine is working but then we don't have original n/a tip so we are using engraving needle this ~~will make~~ making hand dark. so students not using this n/a availability of the tip will be checked

Repaired J. P. P.

Working

Repaired J. P. P. 14/1/22

Repaired J. P. P. 15/2/22

PCB complaint J. P. P. 15/2/22

Repaired J. P. P. 14/3/22

Repaired J. P. P. 15/3/22

Repaired J. P. P. 15/3/22

Repaired J. P. P. 15/3/22

Repaired J. P. P. 14/4/22

6/4/22	P.G.	chair	foot not working	}	}	Dm 28/5/22
18/4/22	U.G.	chair	scissors not working			
	U.G.	chair	movement not working.			
20/4/22	P.G.	X-ray	Not working. no. SNOW 62792. Not file stone remove. Head given to Biju Biomedical. Staff.	Repaired & returned		Fathima 24/5/22 Dm 20/4/22
28/4/22	pu	Chair no: 1100	suction not working	}	}	Dm
"	un	Chair no: 110	motor not working			
29/4/22	pu	motor scissors	not working 05256287			
3/5/22	U.G.	scissors Dicator	Not working	}	}	Fathima 9/5/22
	U.G.	scissors	not working			
	U.G.	Dicator	not working			
12/5/22	pb	welder	not working	Repaired 1+1		Fathima 13/5/22
1/6/22	D.D.	Dicator	Not working LT 800129			

5/2/21	Pa. ① Pa. ② Pa. Pa.	Dental chair ① Dental chair ② Dental Dental chair ④	light complaint, chair movement, ^(foot control) scaler. chair movement, light, chair not working. his rotor - water, switches (chair) scaler.
6/2/21	PG	Dental chair ②	Hand control of chair not working.
12/2/21	PG-6 PG-7	Dental chair "	side switch not working foot control complaint
19/2/21	PG-1 Ontoelav PG	Dental chair Ontoelav	Hand piece water not Not liking ✓
27/2/21	PG-0126	Dental chair	Not working side switch
18/3/21	PG-831	Dental chair	Movement complaint
18/3/21	PG	Autoclave	Not working
8/4/21	PG-131	Dental chair	Arsoles complaint
8/4/21	PG-	Model Tommen	Not water complaint
11/8/21	PG	chair no. 101	Cursing light not working, Suction not working.
11/8/21	PG	cb: no. 1088	Distributor water leaking.
12-8-21	PG	Autoclave	not not working.
13/8/21	PG	chair no. 101	Scaler not measuring.

Repaired Arsoles & Scaler
Complaint Switch complaint (side switch)
Informed to company

Repaired
13/2/21

Repaired
27/2/21

Repaired

Repaired
13/2/21

Repaired

Repaired
27/2/21

L hand complaint to be changed

elbow changed
Repaired
17/3/21

switch to be changed

Repaired
"

side switch to be changed

Repaired
18/3/21

coil complaint to be changed

Repaired
11/3/21

Repaired

Repaired
8/4/21

Repaired

Repaired

Repaired
11/8/21

Repaired

Repaired
10/8/21

under repair

Repaired

Repaired

25/1/21 u.g. 0100 chair suction not working.

28/1/21 pu chair no 20830. bracket water leaking

1/10/21 u.g. chair no 20834 bracket water leaking.

5/10/21 ug Scaled 100/110 ~~at~~ Scaled not working.

28/10/21 pu chair no/0126 water is not coming from printer.

28/10/21 pu chair no. 2081 spray Stryng not working

28/10/21 17110672 ~~at~~ Printer handling not work

23-11-21 P:G. SNO166 X- shaft head not working }
 SNOUX 655 73 No. not file remove }
 Head Given to Balaitha Biomedical staff }

15/12/21 u.g. chair no/0838 Scaled tip water not coming
 water not coming from scales tip

09/1/22 P.G. N56
 TB2
 DBE20220 Hand piece not working

14/1/22 X-ray Room. ~~file~~ RNB Sencer ^{intermittently} stuck

5/2/2022 U.G. Dental chairs (10 No's) All the chair working condition to be checked

SN00X 655 73
 Handpiece returned to dept.

Taken to ston for repair ~~file~~ 2/1/22

All chairs checked and made functional.
 u.g 2nd chair side scotch changed

Reefired

} Reefired

Reefired

Reefired

Received ~~chair~~ ^{chair} 3/1/2022

17/8/21	X-Ray DC Room	Dental chair	light not working. X-Ray cone loose.
25/8/21	pc	Dental chair	Dc/0130 foot control not working
27/8/21	pc	Dental chair	Dc/0126 Suction unit not working.
30/8/21	pc	Dental chair	Dc/0830 Distorter water leaking.
4/9	cu	Dental chair	Dc/0833 Distorter water not coming
16/9	pc	Dental chair no Commutator	Dc/0830 Distorter cutting efficiency low
23/9/21	pc	Dental chair	Dc/0127 foot control not working
23/9/21	cu	Dental chair	Dc/0833 Swing Springs, Distorter not working.
25/9/21	phantom phantom	lab	Water leaking from auster
25/9/21	u/b 0834 0833	u/b	Airster not working water not coming. water leaking. Airster not working water not coming. Air problem. Distorter Leaking.

Repaired	<u>26/8/21</u>
}	<u>20/9/21</u>
	<u>23/9/21</u>
	<u>25/9/21</u>
	<u>Repaired</u>

DATE	CHAIR NO	ITEM	COMPLAINT
23/02/2022	0003	DENTAL CHAIR	Side Switch not working
"	0004	DENTAL CHAIR	NOT Working
"	0006	DENTAL CHAIR	Back Rest NOT working
10/03/2022	0001	DENTAL CHAIR	NOT working
25/04/2022		ICPA MACHINE	NOT working
12/05/2022	0002	DENTAL CHAIR	Light not working
25/05/2022	0003	DENTAL CHAIR	MOVEMENT not working
14/06/2022	0005	Dental Chair	Side switch not working.
23/06/2022	0007	Dental Chair	Back rest not working.
28/06/2022	0006	Dental Chair	Not working.
06/07/2022	0003	Dental Chair	Light not working.
19/07/2022	0001	Dental Chair	Movement not working.
22/07/2022	0005	Dental Chair	Not working.
27/07/2022	0006	Dental Chair	Side switch not working.
30/07/2022	0002	Dental Chair	Light not working.

REMARK OF BIOMEDICAL STAFF	SIGNATURE	REMARK OF DENTIST
Side switch changed	<i>[Signature]</i>	
Main pcb changed	<i>[Signature]</i>	
Motor changed working now		
Repaired		
Send complaint beyond repair. Space not available. X-ray machine from priority clinic shift fixed to one.	<i>[Signature]</i>	
Bulb changed	<i>[Signature]</i>	
Repaired	<i>[Signature]</i>	
Side switch changed	<i>[Signature]</i>	
Repaired	<i>[Signature]</i>	
Repaired	<i>[Signature]</i>	
Motor changed		
Repaired	<i>[Signature]</i>	
Repaired	<i>[Signature]</i>	
Bulb changed working	<i>[Signature]</i>	

DATE	CHAIR NO.	ITEM	COMPLAINT
3/8/2022		Autoclave Persebnice	New pendome used (SANDISK LAB)
8/19/2022	D004	Dental chair	Movement not working
6/08/22	0008	Dental chair	Light Not working
6/05/22	003	Dental chair Dental chair	Movement not working wheel nursing.
	004	chair not working	Movement and rest not working.
1/2/22		Dental Stool (3) Dental Stool	Seat missing Broken Back rest Loose

REMARK OF BIOMEDICAL STAFF	SIGNATURE	REMARK OF
→	<i>[Signature]</i> 2/8/22	
temporarily repaired	<i>[Signature]</i> 6/19/22	
Repair	<i>[Signature]</i> Bar	
Repair	<i>[Signature]</i> Bar	
Repair	<i>[Signature]</i> Bar	
Repair	<i>[Signature]</i> Bar 2/1/22	

24/1/22 X Ray Room & Chair - not working
 17/12 uv cabinet not working
 23/1/2022 - endo uG-

Chair No	Speed	Speed control not working
" 2	-	auto/hi complaint
" 3		OK
" 4	(365)	up/down, hand not working
" 5		OK
" 6		OK
" 7	(368)	not working
" 8		OK
" 9		OK
" 10		not working
" 11		not working
" 12		OK
" 13	(374)	hand piece not working
" 14		not working
" 15	375	up/down, light, hand
" 16	376	up, down not working
" 17		OK
" 18	(346)	OK - Sway spring not working

(Bent)

30/1/22 X Smart endo pt. repair complaint - header loose
 2/8/2022 Antibiotic penicillin New gears (surdisk 16GB)
 11/8/22 endo Surgery. microscope Complaint Speed control

Repaired | *[Signature]*

Side switch changed 4 mm tubings Repaired *[Signature]*

Bulb changed Repaired *[Signature]* 12/18/22

Sway springs changed Repaired *[Signature]* 12/16/22

Sway changed Repaired *[Signature]*

Sway spring changed Repaired *[Signature]* 12/16/22

More (many) for BME acpl: Return back [1-9-22]

[Signature] 2/9/2020

Man return basic 2/8/22